



<div><div>Included Files</div><div>Capture 2022-10-10 15-25-40 Capture 2022-10-10 15-26-21 Capture 2022-10-10 15-26-50</div><div><div>Details</div><div><div>NTA Version:NTA 3.4 Build 3.4.4</div><div>Script Used:SOP Standard Measurement 04-43-57PM 02~</div><div>Time Captured:15:25:30 10/10/2022</div><div>Operator:Ruth</div><div>Pre-treatment:</div><div>Sample Name:S2.3</div><div>Diluent:Cyclohexane</div><div>Remarks:</div></div><div><div>Capture Settings</div><div><div>Camera Type:sCMOS</div><div>Laser Type:Blue405</div><div>Camera Level:12</div><div>Slider Shutter:1200</div><div>Slider Gain:146</div><div>FPS18.7</div><div>Number of Frames:374</div><div>Temperature:22.0 °C</div><div>Viscosity:(Water) 1.0 cP</div><div>Dilution factor:Dilution not recorded</div></div></div><div><div>Analysis Settings</div><div><div>Detect Threshold:14</div><div>Blur Size:Auto</div><div>Max Jump Distance:Auto: 8.9 - 9.4 pix</div></div></div></div></div>	<div><div>Results</div><div><div>Stats: Merged Data</div><div><div>Mean:168.6 nm</div><div>Mode:141.5 nm</div><div>SD:52.2 nm</div><div>D10:120.7 nm</div><div>D50:154.0 nm</div><div>D90:233.4 nm</div></div><div><div>Stats: Mean +/- Standard Error</div><div><div>Mean:167.8 +/- 4.2 nm</div><div>Mode:137.5 +/- 4.0 nm</div><div>SD:51.7 +/- 0.9 nm</div><div>D10:120.0 +/- 5.0 nm</div><div>D50:153.6 +/- 1.5 nm</div><div>D90:226.5 +/- 9.3 nm</div></div><div><div>Concentration:</div><div><div>1.12e+09 +/- 1.15e+08 particles/ml</div><div>61.3 +/- 6.3 particles/frame</div><div>65.5 +/- 5.8 centres/frame</div></div></div></div></div></div>
--	---