


FTLA Concentration / Size graph for Experiment:
Capture 2022-09-26 14-08-51



Averaged FTLA Concentration / Size for Experiment:
Capture 2022-09-26 14-08-51
Error bars indicate + / -1 standard error of the mean

<div><u>Included Files</u> Capture 2022-09-26 14-09-15 Capture 2022-09-26 14-09-52 Capture 2022-09-26 14-10-29 <u>Details</u> NTA Version: NTA 3.4 Build 3.4.4 Script Used: SOP Standard Measurement 04-43-57PM 02~ Time Captured: 14:08:51 26/09/2022 Operator: Ruth Pre-treatment: Sample Name: son. = 15s (2) Diluent: Remarks: <u>Capture Settings</u> Camera Type: sCMOS Laser Type: Blue405 Camera Level: 12 Slider Shutter: 1200 Slider Gain: 146 FPS: 18.7 Number of Frames: 374 Temperature: 22.0 °C Viscosity: (Water) 1.0 cP Dilution factor: Dilution not recorded <u>Analysis Settings</u> Detect Threshold: 24 Blur Size: Auto Max Jump Distance: Auto: 8.7 - 11.0 pix</div>	<div><u>Results</u> Stats: Merged Data Mean: 156.5 nm Mode: 113.3 nm SD: 59.8 nm D10: 102.0 nm D50: 136.6 nm D90: 236.3 nm Stats: Mean +/- Standard Error Mean: 154.6 +/- 19.4 nm Mode: 120.9 +/- 7.2 nm SD: 51.0 +/- 8.8 nm D10: 107.8 +/- 9.2 nm D50: 142.6 +/- 19.5 nm D90: 222.2 +/- 41.0 nm Concentration: 6.29e+08 +/- 4.35e+07 particles/ml 34.4 +/- 2.4 particles/frame 37.1 +/- 2.6 centres/frame</div>
---	---