



<div><div>Included Files</div><div>Capture 2022-10-10 15-02-58 Capture 2022-10-10 15-04-02 Capture 2022-10-10 15-04-36</div><div><div>Details</div><div><div>NTA Version:NTA 3.4 Build 3.4.4</div><div>Script Used:SOP Standard Measurement 04-43-57PM 02~</div><div>Time Captured:14:59:54 10/10/2022</div><div>Operator:Ruth</div><div>Pre-treatment:</div><div>Sample Name:S2.1</div><div>Diluent:Cyclohexane</div><div>Remarks:</div></div><div><div>Capture Settings</div><div><div>Camera Type:sCMOS</div><div>Laser Type:Blue405</div><div>Camera Level:10</div><div>Slider Shutter:696</div><div>Slider Gain:73</div><div>FPS32.3</div><div>Number of Frames:646</div><div>Temperature:22.0 °C</div><div>Viscosity:(Water) 1.0 cP</div><div>Dilution factor:Dilution not recorded</div></div></div><div><div>Analysis Settings</div><div><div>Detect Threshold:15</div><div>Blur Size:Auto</div><div>Max Jump Distance:Auto: 6.8 - 7.1 pix</div></div></div></div></div>	<div><div>Results</div><div><div>Stats: Merged Data</div><div><div>Mean:179.0 nm</div><div>Mode:173.9 nm</div><div>SD:51.6 nm</div><div>D10:128.3 nm</div><div>D50:171.3 nm</div><div>D90:238.9 nm</div></div><div><div>Stats: Mean +/- Standard Error</div><div><div>Mean:181.1 +/- 6.7 nm</div><div>Mode:161.2 +/- 10.0 nm</div><div>SD:52.4 +/- 7.7 nm</div><div>D10:129.0 +/- 1.2 nm</div><div>D50:169.7 +/- 6.2 nm</div><div>D90:241.1 +/- 20.1 nm</div></div><div><div>Concentration:</div><div><div>1.81e+09 +/- 3.36e+08 particles/ml</div><div>99.0 +/- 18.4 particles/frame</div><div>105.1 +/- 19.1 centres/frame</div></div></div></div></div></div>
--	--